
Product Information

Product Name: Auto-Mag® Viral/Pathogen DNA/RNA Isolation Kit	
Product No. AMD-DR05	Lot No. DR0544438
Manufacture Date: 2021-08-30	Expiry Date: 2023-08-30
Note: <i>For Research Use Only. Not for use in diagnostic procedures</i>	

Components in the package:


Description	Storage Condition	Additional Infor.
Auto-Mag® C-5	2-8 °C	
VDR Buffer	15-25 °C	
DRW1 Buffer*	15-25 °C	
DRW2 Buffer*	15-25 °C	
RNA Elution Buffer	15-25 °C	
Proteinase K	2-8 °C	
LEM 1	-20°C	

Quality Control:

Auto-Mag® Viral/Pathogen DNA/RNA Isolation Kit is qualified by isolating Pathogen DNA / RNA from E.coli sample following the protocols outlined in the manual.

Methods/ Specification:	Results
Purified DNA has an A260/280 ratio of ≥ 1.7 .	pass
The functional quality of DNA is evaluated by PCR amplification.	confirm
RNA recovery is analyzed by real-time RT-PCR. The CT value must be < 35 .	pass

Quality authorized by QC:

Signature 

Date: 2021-9-16

ISO Certification:

Manufactured by AMD Biotech Inc. in compliance with ISO 9001-2015 certified quality management system.